


গণপ্রজাতন্ত্রী বাংলাদেশ সরকার
ঔষধ প্রশাসন অধিদপ্তর
ঔষধ ভবন, মহাখালী, ঢাকা-১২১২
www.dgda.gov.bd



বিষয়ঃ ব্যক্তিগত সুরক্ষা সামগ্রী (পিপিই) এবং আনুষঙ্গিক মেডিকেল দ্রব্যাদির স্পেসিফিকেশন প্রসঙ্গে।

বিভিন্ন প্রতিষ্ঠান বর্তমানে দেশে করোনা ভাইরাস আক্রান্ত রোগীদের সেবা প্রদানকারী চিকিৎসক ও অন্যান্য সেবা প্রদানকারীদের চিকিৎসা সেবা প্রদানের সুবিধার্থে বিদেশ থেকে ব্যক্তিগত সুরক্ষা সামগ্রী (পিপিই) আমদানী করছে। এ লক্ষ্যে মানসম্পন্ন ব্যক্তিগত সুরক্ষা সামগ্রী (পিপিই) আমদানী করা আবশ্যিক। কিন্তু লক্ষ্য করা যাচ্ছে যে, কোন কোন ক্ষেত্রে মান সম্মত এ ধরনের সুরক্ষা সামগ্রী আমদানী ও সরবরাহের ক্ষেত্রে ব্যত্যয় ঘটছে।


বিশ্ব স্বাস্থ্য সংস্থা ও প্যান আমেরিকান হেলথ অর্গানাইজেশন (WHO-PAHO) গাইডলাইন বা স্পেসিফিকেশন অনুসরণপূর্বক ব্যক্তিগত সুরক্ষা সামগ্রী (পিপিই) ও এতদসংশ্লিষ্ট আনুষঙ্গিক দ্রব্যাদির আমদানী ও সরবরাহ করা আবশ্যিক।

উপরোক্ত বর্ণনার আলোকে ব্যক্তিগত সুরক্ষা সামগ্রীর (পিপিই) মাননিষ্ঠিতের লক্ষ্যে অত্র অধিদপ্তর কর্তৃক Specification Compliance সনদ গ্রহণপূর্বক আমদানী ও সরবরাহের জন্য সংশ্লিষ্ট সকলকে অনুরোধ করা হল।

WHO PAHO Specification:

Apron	Polyester with PVC coating or 100% PVC or 100% rubber, Waterproof. Minimum basis weight: 250g/m2. Adjustable neck strap (reusable) Covering size: 70-90 cm (width) x 120-150cm (height) or standard adult size.
Goggles (Eye Protector)	Good seal with the skin of the face, Flexible PVC frame to easily fit with all face contours with even pressure, Enclose eyes and the surrounding areas, Accommodate wearers with prescription glasses, Clear plastic lens with fog and scratch resistant treatments, Adjustable band to secure firmly so as not to become loose during clinical activity, Indirect venting to avoid fogging, May be re-usable (provided appropriate arrangements for decontamination are in place) or disposable. EU standard directive 86/686/EEC, EN 166/2002, ANSI/ISEA Z87.1-2010, or equivalent
Face shields (Face Protector)	Made of clear plastic and provides good visibility to both the wearer and the patient. Adjustable band to attach firmly around the head and fit snugly against the forehead, Fog resistant (preferable), completely cover the sides and length of the face, May be re-usable (made of robust material which can be cleaned and disinfected) or disposable. EU standard directive 86/686/EEC, EN166/2002, ANSI/ISEA Z87.1-2010, or equivalent.
Gown	Single use, fluid resistant, disposable, length mid-calf to cover the top of the boots, light colors preferable to better detect possible contamination, thumb/finger loops or elastic cuff to anchor sleeves in place. Option 1: fluid penetration resistant: EN 13795 high performance, or AAMI PB70 level 3 performance or above, or equivalent Option 2: blood borne pathogens penetration resistant: AAMI PB70 level 4 performance, or (EN 14126-B) and partial body protection (EN 13034 or EN 14605), or equivalent
Medical mask	Medical/surgical mask, high fluid resistance, good breathability, internal and external faces should be clearly identified, structured design that does not collapse against the mouth (e.g. duckbill, cup shaped) EN 14683 Type IIR performance ASTM F2100 level 2 or level 3 or equivalent; Fluid resistance at minimum 120 mmHg pressure based on ASTM F1862-07, ISO 22609, or equivalent Breathability: MIL-M-36945C, EN 14683 annex C, or equivalent Filtration efficiency: ASTM F2101, EN14683 annex B, or equivalent
Respirator (N95/FFP2)	"N95" respirator according to US NIOSH, or "FFP2" according to EN 149N95 Good breathability with design that does not collapse against the mouth (e.g. duckbill, cup shaped)
Gloves (Non-Sterile)	Gloves, Examination, Nitrile, powder-Free, Non-Sterile, Cuff Length preferably reach mid-forearm (eg. minimum 280 mm total length, Different sizes EU standard directive 93/42/EEC class1, EN 455, EU standard directive 89/686/EEC category III, EN 374 ANSI/ISEA 105-2011, ASTM d6319-10 or equivalent.
Gloves (Sterile)	Gloves, Surgical, Nitrile, powder-Free, single use. Gloves should have long cuffs, reaching well above the wrist, ideally to mid-forearm. EU standard directive 93/42/EEC class1, EN 455, ANSI/ISEA 105-2011, ASTM 6319-10 or equivalent.

Bags for medical waste	Disposal bag for bio-hazardous waste, 30x50cm, with "Biohazard" print, autoclavable polypropylene. 50 or 70 micra thickness
Body bag	Made of linear enforced, U-shape zipper and 2 zipper pulls with tie ribs. Adult size 250x120cm Protector Body Bag specifications: 6 handles Impermeable, linear reinforced LLDPE, LDPE, EVA, PEVA, (avoid PVC), minimum thickness 400 microns; Should be able to hold 100-125 kilos (200-250 lbs), Should contain no chlorides: burning of chlorides pollute the environment and can cause damage to retort chambers. Body bags should be non-carcinogenic to health of funeral workers when used for cremations. At least 6 handles included in the body bag to allow burial team to hand carry it safely Heat-sealed: insure superior strength and safety, Provide full containment of blood borne pathogens Cracking point of 25 - 32 degrees below zero Shelf life: minimum 10 years , Bag and hands should be of white color


মেজর জেনারেল মোঃ মাহবুবুর রহমান
মহাপরিচালক
ঔষধ প্রশাসন অধিদপ্তর
ফোনঃ ৯৮৮০৮০৩
ই-মেইল: dgda.gov@gmail.com

তারিখ: ০১.০৪.২০২০খ্রিঃ

নং-ডিজিডিএ/২৭-১৩/২০২০/৪৬০

কার্যার্থে অনুলিপি প্রেরণ করা হল (জ্যেষ্ঠতার ক্রমানুসারে নয়)ঃ


- ১। পরিচালক, সেন্ট্রাল মেডিকেল স্টোর ডিপো, স্বাস্থ্য অধিদপ্তর, তেজগাঁও, ঢাকা।
- ২। সভাপতি/মহাসচিব, বাংলাদেশ ঔষধ শিল্প সমিতি, তেজগাঁও, ঢাকা।
- ৩। সভাপতি, ইম্পোর্টার এসোসিয়েশন অব বাংলাদেশ, ঢাকা।
- ৪। সভাপতি/মহাসচিব, মেডিকেল ডিভাইস ইম্পোর্টার এসোসিয়েশন, ঢাকা।
- ৫। সভাপতি, বাংলাদেশ মেডিকেল ইন্সটিটিউট এন্ড হসপিটাল ইকুইপমেন্ট ডিলারস এন্ড ম্যানুফ্যাকচারার্স এসোসিয়েশন, ঢাকা।
- ৬। সভাপতি, বাংলাদেশ সার্জিক্যাল এসোসিয়েশন, ঢাকা।
- ৭। সভাপতি, ডায়াগনস্টিক রিএজেন্ট এন্ড ইকুইপমেন্ট ট্রেডার্স এসোসিয়েশন অব বাংলাদেশ, ঢাকা।

নং-ডিজিডিএ/২৭-১৩/২০২০/৪৬০

তারিখ: ০১.০৪.২০২০খ্রিঃ

অবগতির জন্য অনুলিপি প্রেরণ করা হল (জ্যেষ্ঠতার ক্রমানুসারে নয়)ঃ

- ১। সচিব, স্বাস্থ্য সেবা বিভাগ, স্বাস্থ্য ও পরিবার কল্যাণ মন্ত্রণালয়, বাংলাদেশ সচিবালয়, ঢাকা।
- ২। মহাপরিচালক, সামরিক চিকিৎসা সার্ভিস মহাপরিদপ্তর, ঢাকা ক্যান্টনমেন্ট, ঢাকা।
- ৩। মহাপরিচালক, স্বাস্থ্য অধিদপ্তর, মহাখালী, ঢাকা।
- ৪। অতিরিক্ত মহাপরিচালক (প্রশাসন), স্বাস্থ্য অধিদপ্তর, মহাখালী, ঢাকা।
- ৫। পরিচালক, রোগ নিয়ন্ত্রণ ও লাইন ডাইরেক্টর (সিডিসি), স্বাস্থ্য অধিদপ্তর, মহাখালী, ঢাকা।
- ৬। পরিচালক, হাসপাতাল, স্বাস্থ্য অধিদপ্তর, মহাখালী, ঢাকা।
- ৭। মাননীয় মন্ত্রীর একান্ত সচিব, স্বাস্থ্য ও পরিবার কল্যাণ মন্ত্রণালয়, বাংলাদেশ সচিবালয়, ঢাকা।


মহাপরিচালক
ঔষধ প্রশাসন অধিদপ্তর, ঢাকা।

